Service+ | Implant Prosthetics¹ | Step-by-Step 3shape[>]

Customized Abutment



vodor

Use the suitable scan body (see: Implant Prosthetics Accessories)



Implant selection

Category: Ivoclar Digital Customized Abutment System: Select the desired implant system Kit: Select diameter



Material selection

Material: Service+ Custom Abutment Ti5 Color: polished or unpolished emergence profile Manufacturer: Ivoclar Digital Service+

Hybrid Abutment²



Viteo[®] Base (available from Ivoclar Vivadent)

CAMLOG)



Use the suitable scan body (see: Implant Prosthetics Accessories)



Implant selection Category: Ivoclar Digital Hybrid-Abutment-Crown Select the desired implant system Svstem: Select diameter Kit:



Material selection Select Ivoclar material Material: Select translucency level and shade Color: Manufacturer: Ivoclar Digital Service+

Hybrid Abutment (CAMLOG)





Camlog titanium base (available from

Use the suitable scan body from CAMLOG



Category: Ivoclar Digital Hybrid Abutment (CAMLOG) Select the desired implant system System: Kit: Select diameter



Material selection

Material: Select Ivoclar material Color: Select translucency level and shade Manufacturer: Ivoclar Digital Service+



* Requirements: 3Shape (version 16.1 or higher) Abutment Designer or Implant Bar and Bridge or CAD Points, registration with Ivoclar Digital Service+, import of the library ² also possible with titanium bases from Medentika (2nd generation). Requirements: Installation of the Medentika library and link to Ivoclar materials.

ivoclardigital.com

Service+ | Implant Prosthetics¹ | Step-by-Step 3shape

Implant Bar/Bridge



Use the suitable **scan body** (see: Implant Prosthetics Accessories)



Implant selection

Category: lvoclar Digital Implant Bar/Bridge System: Select the desired implant system Kit: Select diameter



Material selection

3

 Material:
 Service+ Colado® CAD Ti5 Implant or Service+ Colado® CAD CoCr4 Implant

 Color:
 polished or unpolished emergence profile

 Manufacturer:
 Ivoclar Digital Service+

Gingiva Former





Use the suitable **scan body** (see: Implant Prosthetics Accessories)

Abutment	8 6 4 8		
	Service+ Custom Abutment Ti5		
	Category	Ivodar Digital Gingiva former	
	System	Service+ comp. Astra Tech Osse	2
	KR	Ø 3.5/4.0mm (3A-8)-GF	

 Implant selection

 Category:
 Ivoclar Digital Gingiva Former

 System:
 Select the desired implant system

 Kit:
 Select diameter



 Material selection

 Material:
 Service+ Custom Abutment Ti5

 Color:
 polished or unpolished emergence profile

 Manufacturer:
 Ivoclar Digital Service+

Hybrid Bridge³



Hybrid bridge



Camlog titanium base³ Use the suitable scan body from CAMLOG



Implant selectionCategory:Ivoclar Digital Hybrid BridgeSystem:Select the desired implant systemKit:Select diameter



Material selection

 Material:
 Select Ivoclar material

 Color:
 Select translucency level and shade

 Manufacturer:
 Ivoclar Digital Service+



^a also possible with titanium bases from Medentika (2rd generation). Requirements: Installation of the Medentika library and link to Ivoclar materials.

Date information prepared: 2018-01-11, Rev. 1

ivoclardigital.com